



December 08

Pathogen Solutions Limited – Medixair

Basic research and development has been completed and the product may have potential value; in use evaluations/trials are now needed in an healthcare setting. (recommendation 2)

This product is an air purification device that recirculates air after exposure to UV light. The submission included two studies, both performed in the same hospital. One looked at the effect of the device on environmental contamination with MRSA, the second looked at the effect on Clostridium difficile infection rates. Further data is needed to confirm whether the demonstrated reductions were due to use of this device or to other factors.