



June 2010

Tristel Solutions Ltd– Plasmair T-2006

A potentially useful new concept but insufficiently validated; more research and development is required before it is ready for evaluation in practice (recommendation 3)

This product is a mobile, stand-alone air decontamination unit.

More trials are required to demonstrate sufficient air movement through the air cleaning system. The link between air contamination and infection transmission in general hospital areas is not established.